



EU Type Examination Certificate No CH-MI004-08007-05

| | |
|---------------------------------|---|
| <i>Applicant:</i> | Coster Group s.r.l. Divisione Tecnica Via San G.B.De La Salle 4/A 20132 Milano (MI) Italy |
| <i>Requirements:</i> | Ordinance of 15 February 2006 on Measuring Instruments (SR 941.210) and Ordinance of the FDJP of 19 March 2006 on Heat Meters (SR 941.231); Directive 2014/32/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments (MID) and the instrument-specific annex VI (MI-004) |
| <i>Conformity standards:</i> | OIML R75:2002, EN 1434-1:2007 |
| <i>Type of instrument:</i> | Calculator as a subassembly of a heat meter |
| <i>Type designation:</i> | IET 7383, IET 7343, IET 7183, IET 7143 |
| <i>Accuracy class:</i> | 2 (according to OIML R75) |
| <i>Characteristics:</i> | Min. temperature ϑ_{min} : 0 °C Max. temperature ϑ_{max} : 170 °C Min. temperature diff. $\Delta\vartheta_{min}$: 3 K Max. temperature diff. $\Delta\vartheta_{max}$: 150 K Environmental class: A, 5 °C ... 45 °C Protection class: IP54 Mechanical class: M1 |
| <i>Certificate valid until:</i> | 19 February 2021 |
| <i>Notified body</i> | Conformity Evaluation Body METAS-Cert No. 1259 3003 Berne-Wabern, 20 February 2019 |
| <i>Approved by</i> | Gulian Couvreur, Head of sector METAS-Cert |

